



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

### ***Advanced Metrology Est.***

***Khalid Ibn Alwaleed Street, Alrasah Alshamaliah, Al Khobar, 31952, Saudi Arabia***

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

### **ISO/IEC 17025:2005**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

***Chemical, Dimensional, Electrical, Mass, Force, & Weighing Devices, Mechanical, Thermodynamic, and Time & Frequency Calibration***  
*(As detailed in the supplement)*

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen  
President/Operations Manager

*Initial Accreditation Date:*

December 1, 2017

*Issue Date:*

December 1, 2017

*Expiration Date:*

March 31, 2020

*Accreditation No.:*

90462

*Certificate No.:*

L17-528

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjllabs.com](http://www.pjllabs.com)*



# Certificate of Accreditation: Supplement

## Advanced Metrology Est.

Khalid Ibn Alwaleed Street, Alrahah Alshamaliah, Al Khobar, 31952, Saudi Arabia  
 Contact Name: Dr. Mohammad Hourani Phone: 966-13-833-1203

Accreditation is granted to the facility to perform the following calibrations:

### Chemical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|--|
| Conductivity Meter <sup>FO</sup>       | 100 $\mu$ S                                 | 0.57 $\mu$ S   | Conductivity Solutions                             |
|  | 10 00 $\mu$ S                               | 4.2 $\mu$ S  |  |
|  | 10 000 $\mu$ S                              | 340 $\mu$ S  |  |
| pH Meter/Probe <sup>FO</sup>           | 4 pH  | 0.032 pH   | Buffer Solutions                                   |
|  | 7 pH  |  |  |
|  | 10 pH                                       |  |  |

### Dimensional

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED                                 |
|--|---|--|--|
| Calipers <sup>F</sup>                  | 1 mm to 100 mm                              | 0.006 2 mm   | Gage Blocks, DIN 862, DIN 878, DIN 879, ISO  |
|  | 100 mm to 300 mm                            | 0.007 2 mm   |  |
|  | 300 mm to 600 mm                            | 0.009 1 mm   |  |
| Dial/Thickness Gages <sup>F</sup>      | 1 mm to 100 mm                              | 0.006 2 mm   | Universal Measuring Machine, EURAMET cg 6  |
|  | 100 mm to 300 mm                            | 0.008 5 mm   |  |
| Diameter Inside <sup>F</sup>           | 0.01 mm to 300 mm                           | 0.001 6 mm   | Universal Measuring Machine, EURAMET cg 6  |
| Diameter Outside <sup>F</sup>          | 25 mm to 300 mm                             | 0.008 4 mm   |  |
|  | 300 mm to 1 000 mm                          | 0.57 mm  |  |
| Gage Blocks <sup>F</sup>               | 0.000 5 mm to 100 mm                        | $(4.45 \times 10^{-4} + 5 \times 10^{-6}L)$ mm                               | Gage Block Comparator and master Gage Clocks Grade 0, The Gage Block Handbook-NIST |
| Height Gage <sup>F</sup>               | 0 mm to 600 mm                              | $(6.35 \times 10^{-3} + 5 \times 10^{-6}L)$ mm                               | Gage Block Set   |
| Outside Micrometers <sup>F</sup>       | 1 mm to 100 mm                              | 0.006 2 mm   | Gage Blocks, DIN 862, DIN 878, DIN 879, ISO  |
|  | 100 mm to 300 mm                            | 0.007 2 mm   |  |
|  | 300 mm to 600 mm                            | 0.009 1 mm   |  |
| Protractors <sup>F</sup>               | 0° to 90°                                   | 0.4°   | Gage Blocks/Sine Bar   |



# Certificate of Accreditation: Supplement

## Advanced Metrology Est.

Khalid Ibn Alwaleed Street, Alrakah Alshamaliah, Al Khobar, 31952, Saudi Arabia  
 Contact Name: Dr. Mohammad Hourani Phone: 966-13-833-1203

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

| MEASURED INSTRUMENT,<br>QUANTITY OR GAUGE                                   | RANGE OR NOMINAL<br>DEVICE SIZE AS<br>APPROPRIATE | CALIBRATION AND<br>MEASUREMENT<br>CAPABILITY EXPRESSED<br>AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION<br>EQUIPMENT<br>AND REFERENCE<br>STANDARDS USED                                    |
|---|---|---|--|
| Equipment to Measure AC Voltage <sup>F</sup><br>(at the listed frequencies) |   |   | Transmille Precision<br>Multi-Product Calibrator<br>3041A, Multi-Function<br>Workstation EA015 |
| 10 Hz to 206 Hz   | 0 mV to 200 mV                                    | 0.1 mV  |  |
| 206 Hz to 10 kHz  | 0 mV to 200 mV                                    | 0.1 mV  |  |
| 10 kHz to 100 kHz   | 0 mV to 200 mV                                    | 2.5 mV  |  |
| Equipment to Measure AC Voltage <sup>F</sup><br>(at the listed frequencies) |   |   |  |
| 10 Hz to 206 Hz   | 200 mV to 2 V                                     | 1 mV  |  |
| 206 Hz to 10 kHz  | 200 mV to 2 V                                     | 1 mV  |  |
| 10 kHz to 100 kHz   | 200 mV to 2 V                                     | 2.3 mV  |  |
| Equipment to Measure AC Voltage <sup>F</sup><br>(at the listed frequencies) |   |   |  |
| 10 Hz to 206 Hz   | 2 V to 20 V                                       | 10 mV   |  |
| 206 Hz to 10 kHz  | 2 V to 20 V                                       | 6 mV  |  |
| 10 kHz to 100 kHz   | 2 V to 20 V                                       | 20 mV   |  |
| Equipment to Measure AC Voltage <sup>F</sup><br>(at the listed frequencies) |   |   |  |
| 10 Hz to 206 Hz   | 20 V to 200 V                                     | 150 mV  |  |
| 206 Hz to 10 kHz  | 20 V to 200 V                                     | 180 mV  |  |
| 10 kHz to 100 kHz   | 20 V to 200 V                                     | 300 mV  |  |
| Equipment to Measure AC Voltage <sup>F</sup><br>(at the listed frequencies) |   |   |  |
| 10 Hz to 206 Hz   | 200 V to 1 kV                                     | 480 mV  |  |
| 206 Hz to 10 kHz  | 200 V to 1 kV                                     | 440 mV  |  |
| 10 kHz to 100 kHz   | 200 V to 1 kV                                     | 580 mV  |  |
| Equipment to Measure AC Voltage <sup>F</sup><br>(at the listed frequencies) |   |   |  |
| 10 Hz to 206 Hz   | 0 $\mu$ A to 200 $\mu$ A                          | 0.29 $\mu$ A  |  |
| 206 Hz to 10 kHz  | 0 $\mu$ A to 200 $\mu$ A                          | 0.28 $\mu$ A  |  |
| Equipment to Measure AC Voltage <sup>F</sup><br>(at the listed frequencies) |   |   |  |
| 10Hz to 206Hz   | 200 $\mu$ A to 2 mA                               | 3 $\mu$ A   |  |
| 206Hz to 10kHz  | 200 $\mu$ A to 2 mA                               | 2 $\mu$ A   |  |
| Equipment to Measure AC Voltage <sup>F</sup><br>(at the listed frequencies) |   |   |  |
| 10Hz to 206Hz   | 2 mA to 20 mA                                     | 0.03 mA   |  |
| 206Hz to 10kHz  | 2 mA to 20 mA                                     | 0.02 mA   |  |



# Certificate of Accreditation: Supplement

## Advanced Metrology Est.

Khalid Ibn Alwaleed Street, Alrakah Alshamaliah, Al Khobar, 31952, Saudi Arabia  
 Contact Name: Dr. Mohammad Hourani Phone: 966-13-833-1203

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

| MEASURED INSTRUMENT,<br>QUANTITY OR GAUGE                                   | RANGE OR NOMINAL<br>DEVICE SIZE AS<br>APPROPRIATE | CALIBRATION AND<br>MEASUREMENT<br>CAPABILITY EXPRESSED<br>AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION<br>EQUIPMENT<br>AND REFERENCE<br>STANDARDS USED                                    |
|---|---|---|--|
| Equipment to Measure AC Voltage <sup>F</sup><br>(at the listed frequencies) |   |   | Transmille Precision<br>Multi-Product Calibrator<br>3041A, Multi-Function<br>Workstation EA015 |
| 10 Hz to 206 Hz   | 20 mA to 200 mA                                   | 0.29 mA   |  |
| 206 Hz to 10 kHz  | 20 mA to 200 mA                                   | 0.22 mA   |  |
| Equipment to Measure AC Voltage <sup>F</sup><br>(at the listed frequencies) |   |   |  |
| 10 Hz to 206 Hz   | 200 mA to 2 A                                     | 2.9 mA  |  |
| 206 Hz to 10 kHz  | 200 mA to 2 A                                     | 2.6 mA  |  |
| Equipment to Measure AC Voltage <sup>F</sup><br>(at the listed frequencies) |   |   |  |
| 10 Hz to 206 Hz   | 2 A to 30 A                                       | 28 mA   |  |
| 206 Hz to 10 kHz  | 2 A to 30 A                                       | 16 mA   |  |
| Equipment to Measure AC Voltage <sup>F</sup><br>(at the listed frequencies) |   |   |  |
| 10 Hz to 206 Hz   | 30 A to 60 A                                      | 24 mA   |  |
| 206 Hz to 10 kHz  | 30 A to 60 A                                      | 140 mA  |  |
| Equipment to Measure AC Voltage <sup>F</sup><br>(at the listed frequencies) |   |   |  |
| 10 Hz to 206 Hz   | 60 A to 300 A                                     | 280 mA  |  |
| 206 Hz to 10 kHz  | 60 A to 300 A                                     | 200 mA  |  |
| Equipment to Measure AC Voltage <sup>F</sup><br>(at the listed frequencies) |   |   |  |
| 10 Hz to 206 Hz   | 300 A to 1 500 A                                  | 4.1 A   |  |
| 206 Hz to 10 kHz  | 300 A to 1 500 A                                  | 2.9 A   |  |
| Equipment to Output AC Voltage <sup>FO</sup><br>(at the listed frequencies) |   |   | Fluke 8845A and<br>Fluke 376   |
| 3 kHz to 20kHz  | 0 mV to 100 mV                                    | 0.02 mV/V + 0.06 mV   |  |
| 20 kHz to 100 kHz   | 0 mV to 100 mV                                    | 0.01 mV/V + 0.16 mV   |  |
| 100 kHz to 300 kHz  | 0 mV to 100 mV                                    | 0.08 mV/V + 1 mV  |  |
| Equipment to Output AC Voltage <sup>FO</sup><br>(at the listed frequencies) |   |   |  |
| 3 kHz to 20 kHz   | 0.1 V to 1 V                                      | 0.02 mV/V + 0.06 mV   |  |
| 20 kHz to 100 kHz   | 0.1 V to 1 V                                      | 0.01 mV/V + 0.16 mV   |  |
| 100 kHz to 300 kHz  | 0.1 V to 1 V                                      | 0.08 mV/V + 1 mV  |  |



# Certificate of Accreditation: Supplement

## Advanced Metrology Est.

Khalid Ibn Alwaleed Street, Alrakah Alshamaliah, Al Khobar, 31952, Saudi Arabia  
 Contact Name: Dr. Mohammad Hourani Phone: 966-13-833-1203

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

| MEASURED INSTRUMENT,<br>QUANTITY OR GAUGE                                   | RANGE OR NOMINAL<br>DEVICE SIZE AS<br>APPROPRIATE | CALIBRATION AND<br>MEASUREMENT<br>CAPABILITY EXPRESSED<br>AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION<br>EQUIPMENT<br>AND REFERENCE<br>STANDARDS USED |
|---|---|---|---|
| Equipment to Output AC Voltage <sup>FO</sup><br>(at the listed frequencies) |   |   | Fluke 8845A and<br>Fluke 376                                |
| 3 kHz to 20 kHz   | 1 V to 10 V                                       | 0.02 V/V + 0.06 V   |   |
| 20 kHz to 100 kHz   | 1 V to 10 V                                       | 0.01 V/V + 0.16 V   |   |
| 100 kHz to 300 kHz  | 1 V to 10 V                                       | 0.08 V/V + 1 V  |   |
| Equipment to Output AC Voltage <sup>FO</sup><br>(at the listed frequencies) |   |   |   |
| 3 kHz to 20 kHz   | 10 V to 100 V                                     | 0.02 V/V + 0.06 V   |   |
| 20 kHz to 100 kHz   | 10 V to 100 V                                     | 0.01 V/V + 0.16 V   |   |
| 100 kHz to 300 kHz  | 10 V to 100 V                                     | 0.08 V/V + 1 V  |   |
| Equipment to Output AC Voltage <sup>FO</sup><br>(at the listed frequencies) |   |   |   |
| 3 kHz to 20kHz  | 100 V to 750 V                                    | 0.02 V/V + 0.06 V   |   |
| 20 kHz to 100 kHz   | 100 V to 750 V                                    | 0.01 V/V + 0.16 V   |   |
| 100 kHz to 300 kHz  | 100 V to 750 V                                    | 0.08 V/V + 1 V  |   |
| Equipment to Output AC Current <sup>FO</sup><br>(at the listed frequencies) |   |   |   |
| 10 Hz to 5 kHz  | 0 mA to 10 mA                                     | 2 $\mu$ A/A + 8 $\mu$ A   |   |
| 5 kHz to 10 kHz   | 0 mA to 10 mA                                     | 4 $\mu$ A/A + 25 $\mu$ A  |   |
| Equipment to Output AC Current <sup>FO</sup><br>(at the listed frequencies) |   |   |   |
| 10 Hz to 5 kHz  | 10 mA to 100 mA                                   | 2 $\mu$ A/A + 80 $\mu$ A  |   |
| 5 kHz to 10 kHz   | 10 mA to 100 mA                                   | 4 $\mu$ A/A + 250 $\mu$ A   |   |
| Equipment to Output AC Current <sup>FO</sup><br>(at the listed frequencies) |   |   |   |
| 10 Hz to 5 kHz  | 0.1 A to 1 A                                      | 2 mA/A + 0.8 mA   |   |
| 5 kHz to 10 kHz   | 0.1 A to 1 A                                      | 4 mA/A + 5 mA   |   |
| Equipment to Output AC Current <sup>FO</sup><br>(at the listed frequencies) |   |   |   |
| 10 Hz to 5 kHz  | 1 A to 3 A  | 2 mA/A + 2.4 mA   |   |
| 5 kHz to 10 kHz   | 1 A to 3 A  | 4 mA/A + 15 mA  |   |
| Equipment to Output AC Current <sup>FO</sup><br>(at the listed frequencies) |   |   |   |
| 10 Hz to 5 kHz  | 3 A to 10 A                                       | 2 mA/A + 8 mA   |   |
| 5 kHz to 10 kHz   | 3 A to 10 A                                       | 4 mA/A + 50 mA  |   |



# Certificate of Accreditation: Supplement

## Advanced Metrology Est.

Khalid Ibn Alwaleed Street, Alrakah Alshamaliah, Al Khobar, 31952, Saudi Arabia  
 Contact Name: Dr. Mohammad Hourani Phone: 966-13-833-1203

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE                                      | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED                                    |
|---|---|--|---|
| Equipment to Output AC Current <sup>FO</sup><br>(at the listed frequencies) |   |  | Fluke 8845A and Fluke 376   |
| 5 Hz to 500 Hz  | 10 A to 1 000 A                             | 4 % of Reading   |   |
| 5 Hz to 500 Hz  | 1 000 A to 2 500 A                          | 5 % of Reading   |   |
| Equipment to Measure DC Voltage <sup>F</sup>                                | 200 mV                                      | 1.5 $\mu$ V  | Transmille Precision Multi-Product Calibrator 3041A, Multi-Function Workstation EA015 |
|   | 2 V   | 3.5 $\mu$ V  |   |
|   | 20 V  | 100 $\mu$ V  |   |
|   | 200 V                                       | 1 mV   |   |
|   | 1 kV  | 10 mV  |   |
| Equipment to Measure DC Current <sup>F</sup>                                | 200 $\mu$ A                                 | 5 nA   |   |
|   | 2 mA  | 100 nA   |   |
|   | 20 mA                                       | 400 nA   |   |
|   | 200 mA                                      | 6 $\mu$ A  |   |
|   | 2 A   | 100 $\mu$ A  |   |
|   | 30 A  | 10 mA  |   |
|   | 60 A  | 35 mA  |   |
|   | 300 A                                       | 35 mA  |   |
|   | 1 500 A                                     | 35 mA  |   |
| Equipment to Measure Resistance <sup>F</sup>                                | 1 $\Omega$                                  | 0.8 m $\Omega$   |   |
|   | 10 $\Omega$                                 | 1 m $\Omega$   |   |
|   | 100 $\Omega$                                | 1.3 m $\Omega$   |   |
|   | 1 k $\Omega$                                | 4 m $\Omega$   |   |
|   | 10 k $\Omega$                               | 21 m $\Omega$  |   |
|   | 100 k $\Omega$                              | 620 m $\Omega$   |   |
|   | 1 M $\Omega$                                | 14 $\Omega$  |   |
|   | 10 M $\Omega$                               | 150 $\Omega$   |   |
|   | 100 M $\Omega$                              | 5 k $\Omega$   |   |
| Equipment to Output Capacitance <sup>F</sup>                                | 1 $\mu$ F                                   | 0.000 25 $\mu$ F   |   |
|   | 10 $\mu$ F                                  | 23 $\mu$ F   |   |
|   | 100 $\mu$ F                                 | 0.085 $\mu$ F  |   |
|   | 1 000 $\mu$ F                               | 3.3 $\mu$ F  |   |
|   | 20 nF                                       | 0.04 nF  |   |
|   | 50 nF                                       | 0.1 nF   |   |
|   | 100 nF                                      | 0.1 nF   |   |
|   | 1 000 nF                                    | 0.25 nF  |   |



# Certificate of Accreditation: Supplement

## Advanced Metrology Est.

Khalid Ibn Alwaleed Street, Alrakah Alshamaliah, Al Khobar, 31952, Saudi Arabia  
 Contact Name: Dr. Mohammad Hourani Phone: 966-13-833-1203

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE       | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED                                    |
|--|---|--|---|
| Equipment to Measure Frequency <sup>F</sup>  | 100 Hz                                      | 0.000 02 Hz  | Transmille Precision Multi-Product Calibrator 3041A, Multi-Function Workstation EA015 |
|  | 1 000 Hz                                    | 0.000 2 Hz   |   |
|  | 1 kHz                                       | 0.000 002 kHz  |   |
|  | 10 kHz                                      | 0.000 002 kHz  |   |
|  | 100 kHz                                     | 0.000 02 kHz   |   |
|  | 10 MHz                                      | 20 Hz  |   |
|  | 300 MHz                                     | 600 Hz   |   |
|  | 600 MHz                                     | 1.2 kHz  |   |
| Equipment to Output DC Voltage <sup>FO</sup> | 0 mV to 100 mV                              | 0.1 $\mu$ V/V + 7 $\mu$ V  | Fluke 8845A and Fluke 376   |
|  | 0.1 V to 1 V                                | 0.08 mV/V + 0.07 mV  |   |
|  | 1 V to 10 V                                 | 0.07 mV/V + 0.14 mV  |   |
|  | 10 V to 100 V                               | 0.09 mV/V + 1 mV   |   |
|  | 100 V to 1000 V                             | 0.09 mV/V + 12 mV  |   |
| Equipment to Output DC Current <sup>FO</sup> | 0 $\mu$ A to 100 $\mu$ A                    | 0.001 $\mu$ A/A + 0.05 $\mu$ A   |   |
|  | 0.1 mA to 1 mA                              | 0.001 mA/A + 0.5 $\mu$ A   |   |
|  | 1 mA to 10 mA                               | 0.001 mA/A + 5 $\mu$ A   |   |
|  | 10 mA to 100 mA                             | 0.01 mA/A + 50 $\mu$ A   |   |
|  | 0.1 A to 1 A                                | 1 mA/A + 0.5 mA  |   |
|  | 1 A to 3 A                                  | 1 mA/A + 1.5 mA  |   |
|  | 3 A to 10 A                                 | 1 mA/A + 5 mA  |   |
|  | 10 A to 1 000 A                             | 4 % of Reading   |   |
| Equipment to Measure Frequency <sup>FO</sup> | 3 Hz to 5 Hz                                | 0.2 % of Reading   |   |
|  | 5 Hz to 10 Hz                               | 0.1 % of Reading   |   |
|  | 10 Hz to 40 Hz                              | 0.06 % of Reading  |   |
|  | 40 Hz to 300 Hz                             | 0.02 % of Reading  |   |
| Equipment to Output Resistance <sup>FO</sup> | 0 $\Omega$ to 100 $\Omega$                  | 0.2 m $\Omega$ / $\Omega$ + 8 m $\Omega$                                     |   |
|  | 100 $\Omega$ to 1 k $\Omega$                | 2 $\Omega$ / $\Omega$ + 0.08 $\Omega$  |   |
|  | 1 k $\Omega$ to 10 k $\Omega$               | 2 $\Omega$ / $\Omega$ + 8 $\Omega$   |   |
|  | 10 k $\Omega$ to 100 k $\Omega$             | 2 $\Omega$ / $\Omega$ + 80 $\Omega$  |   |
|  | 100 k $\Omega$ to 1 M $\Omega$              | 2 k $\Omega$ / $\Omega$ + 0.08 k $\Omega$                                    |   |
|  | 1 M $\Omega$ to 10 M $\Omega$               | 2 k $\Omega$ / $\Omega$ + 8 k $\Omega$                                       |   |
|  | 10 M $\Omega$ to 100 M $\Omega$             | 16 k $\Omega$ / $\Omega$ + 20 k $\Omega$                                     |   |



# Certificate of Accreditation: Supplement

## Advanced Metrology Est.

Khalid Ibn Alwaleed Street, Alrakah Alshamaliah, Al Khobar, 31952, Saudi Arabia  
 Contact Name: Dr. Mohammad Hourani Phone: 966-13-833-1203

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE              | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE    | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED                                    |
|---|--|--|---|
| Temperature Indicators and Simulators <sup>FO</sup> | -200 °C to 1 750 °C                            | 0.4 °C   | Temperature Calibrator LR-Cal model LTC 100 EURAMET cg 11 V3                          |
| Temperature Sources / Blocks <sup>FO</sup>          | -200 °C to 1 370 °<br>(-5.891 mV to 54.807 mV) | 0.4 °C   |   |
| Thermocouples <sup>FO</sup>                         | -200 °C to 1 200 °C<br>(-7.89 mV to 69.539 mV) | 0.4 °C   | Temperature Calibrator LR-Cal model LTC 100 EURAMET cg 8 V2                           |
| Inductance <sup>F</sup>                             | 10 H   | 6 mH   | Transmille Precision Multi-Product Calibrator 3041A, Multi-Function Workstation EA015 |
| Phase Angle Full Range <sup>F</sup>                 | 90°  | 0.5°   |   |
| Power Consumption Energy <sup>FO</sup>              | 750 kWh  | 0.001 5 kWh  |   |
| Reactive Power <sup>FO</sup>                        | 750 kVAr                                       | 0.000 26 kVAr  |   |
| Time Base <sup>F</sup>                              | 2 Hz   | 0.000 1 Hz   |   |

### Mass, Force, & Weighing Devices

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED  |
|--|---|--|---|
| Analytical Balances <sup>FO</sup>      | 1 mg to 320 g                               | $(2.1 \times 10^{-3} + 2.1 \times 10^{-5}Wt)$                                | Class E2 Weights OIML-R E76-1   |
| Electronic Balances <sup>FO</sup>      | 320 g to 20 kg                              | $(3.96 \times 10^{-3} + 1.48 \times 10^{-2}Wt)$                              | Class F1 Weights IML-R E76-1  |
| Weighing Balances <sup>FO</sup>        | 20 kg to 3 000 kg                           | $(2.99 \times 10^{-1} + 5.1 \times 10^{-5}Wt)$                               | Class M1 Weights OIML-R E76-1   |
| Weights and Weight Sets <sup>F</sup>   | 1 mg  | 0.016 mg   | Double Substitution with Air Buoyancy Correction, Class E2 Mass Set, Mass Comparators, OIML 111-1 |
|  | 2 mg  | 0.016 mg   |   |
|  | 5 mg  | 0.016 mg   |   |
|  | 10 mg                                       | 0.016 mg   |   |
|  | 20 mg                                       | 0.016 mg   |   |
|  | 50 mg                                       | 0.017 mg   |   |
|  | 100 mg                                      | 0.017 mg   |   |
|  | 200 mg                                      | 0.017 mg   |   |
|  | 500 mg                                      | 0.017 mg   |   |
|  | 1 g   | 0.018 mg   |   |
|  | 2 g   | 0.02 mg  |   |
|  | 5 g   | 0.02 mg  |   |
|  | 10 g  | 0.025 mg   |   |
| 20 g                                   | 0.03 mg                                     |  |   |





# Certificate of Accreditation: Supplement

## Advanced Metrology Est.

Khalid Ibn Alwaleed Street, Alrakah Alshamaliah, Al Khobar, 31952, Saudi Arabia  
 Contact Name: Dr. Mohammad Hourani Phone: 966-13-833-1203

Accreditation is granted to the facility to perform the following calibrations:

### Mass, Force, & Weighing Devices

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED  |
|--|---|--|---|
| Weights and Weight Sets <sup>F</sup>   | 50 g  | 0.034 mg   | Double Substitution with Air Buoyancy Correction, Class E2 Mass Set, Mass Comparators, OIML 111-1 |
|  | 100 g                                       | 0.053 mg   |   |
|  | 200 g                                       | 0.11 mg  |   |
|  | 500 g                                       | 0.26 mg  |   |
|  | 1 kg  | 0.51 mg  |   |
|  | 2 kg  | 1.8 mg   |   |
|  | 5 kg  | 3 mg   |   |
|  | 10 kg                                       | 9.5 mg   |   |
|  | 20 kg                                       | 18 mg  |   |

### Mechanical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE           | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED         |
|--|---|--|--|
| Pressure Gages / Switches / Valves <sup>FO</sup> | 0 psi to 100 psi                            | 0.022 psi  | Master Pressure Gages with pumps, DKD R61E                 |
|  | 100 psi to 1 000 psi                        | 0.19 psi   |  |
|  | 1 000 psi to 10 000 psi                     | 1.0 psi  |  |
|  | 10 000 psi to 36 000 psi                    | 5 psi  |  |
| Tension and Compression Machines <sup>O</sup>    | 0 kg to 500 kg                              | 0.29 kg  | Load Cell or Standard Weights                              |
|  | 5 kN to 2 000 kN                            | 2.3 kN   |  |
| Torque Testers <sup>F</sup>                      | 0 Nm to 25 Nm                               | 0.31 Nm  | Torque Arm EURAMET cg14                                    |
|  | 25 Nm to 1 500 Nm                           | 0.31 Nm  |  |
| Torque Wrench <sup>F</sup>                       | 0 Nm to 25 Nm                               | 0.31 Nm  | Torque Testing ISO 6789                                    |
|  | 25 Nm to 400 Nm                             | 0.31 Nm  |  |
|  | 400 Nm to 1 500 Nm                          | 0.31 Nm  |  |
| Vacuum Gage <sup>FO</sup>                        | -0.95 bar to 0 bar                          | 0.058 psi  | Master Vacuum Gage and Vacuum Pump DKD R61E                |
| Volumetric Flasks <sup>F</sup>                   | 0 L to 20 L                                 | $(2 \times 10^{-1} + 4.9 \times 10^{-4}V)$ L                                 | Weight Set and Balance, OIML R43, OIML R120, EURAMET cg 19 |
| Volumetric Meters <sup>O</sup>                   | 0 L to 100 L                                | $(1.95 \times 10^{-1} + 4.99 \times 10^{-3}V)$ L                             | Stainless Steel cans, OIML R120, EURAMET cg 19             |



# Certificate of Accreditation: Supplement

## Advanced Metrology Est.

Khalid Ibn Alwaleed Street, Alrakah Alshamaliah, Al Khobar, 31952, Saudi Arabia  
Contact Name: Dr. Mohammad Hourani Phone: 966-13-833-1203

Accreditation is granted to the facility to perform the following calibrations:

### Thermodynamic

| MEASURED INSTRUMENT, QUANTITY OR GAUGE       | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED  |
|--|---|--|---|
| Humidity Measuring Instruments <sup>FO</sup> | 10 % to 95 %                                | 0.51 % RH  | Kambic Chamber<br>PKK50 with Rotronic<br>HygroLog HL-20D  |
| IR Thermometers <sup>FO</sup>                | -40 °C to 130 °C                            | 0.34 °C  | Temperature dry block<br>LR-Cal model PULSAR<br>35CU-2l , Kambic oil<br>bath model OBM-LT and<br>4 wire platinum resistance<br>RTD(PT100)<br>ASTM E77<br>ASTM E563<br>ASTM E1 |
|  | 130 °C to 600 °C                            | 0.44 °C  |   |
| RTDs (PT100) <sup>FO</sup>                   | -40 °C to 130 °C                            | 0.46 °C  |   |
|  | 130 °C to 600 °C                            | 0.53 °C  |   |
| Temperature Gage / Thermometer <sup>FO</sup> | -40 °C to 130 °C                            | 0.46 °C  |   |
|  | 130 °C to 600 °C                            | 0.53 °C  |   |

### Time & Frequency

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|--|
| Sound Level Meter <sup>F</sup>         | 94 dB                                       | 0.65 dB  | Sound Level Calibrator<br>REED SC-05               |
|  | 114 dB                                      |  |  |

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor  $k$  (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer<sup>F</sup> would mean that the laboratory performs this calibration at its fixed location.
4. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations. Example: Outside Micrometer<sup>O</sup> would mean that the laboratory performs this calibration onsite at the customer's location.



## *Certificate of Accreditation: Supplement*

### **Advanced Metrology Est.**

Khalid Ibn Alwaleed Street, Alrahah Alshamaliah, Al Khobar, 31952, Saudi Arabia  
Contact Name: Dr. Mohammad Hourani Phone: 966-13-833-1203

*Accreditation is granted to the facility to perform the following calibrations:*

5. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer<sup>FO</sup> would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
6. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
7. The term L represents length in inches or millimeters as appropriate to the uncertainty statement.
8. The term Wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.
9. The term V represents volume as appropriate to the uncertainty statement.

